

Certificate

Certificate no.**CU 72405579 0001****License Holder:**

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Manufacturing Plant:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Report Number:

DE24T27E 001

Client Reference:

Mr. Steffen Geißler

Certification acc. to:

UL 61010-1:2012 R6.23
CSA C22.2 No. 61010-1-12 (R2022) R6.23

Product Information**Certified Product:**

Auto-sampler

Model Designation:

AS 60

Technical Data:

Rated voltage : 24 V DC
Rated current : 2.5 A
Rated power : 30 W
Protection class: III

Remarks:**Appendix:**

1, 1-3

Date of issue:

2024-10-08
(yr/mo/day)



TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 **TÜVRheinland®**