

Certificate



Certificate no.

CU 72300095 01

License Holder:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Manufacturing Plant:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Test report no.: USA- 31982225 008

Client Reference:

Tested to: UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
UL 61010-2-061:2019
CSA C22.2 No. 61010-2-061:19

Certified Product: ICP Emission Spectrometer (ICP-OES)

License Fee - Units

Model Designation:
PlasmaQuant 9100, PlasmaQuant 9100 Elite

7

Rating: 230V AC, 50/60Hz, 4500VA

Replaces certificate no.: CU 72211467.01, CU 72211467.02

Appendix: 1, 1-12

7

Licensed Test mark:

Date of Issue
(day/mo/yr)
17/07/2023

