

Certificate

Certificate no.**CU 72405461 0001****License Holder:**

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Manufacturing Plant:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Report Number: DE243YT8 001**Client Reference:** Mr. Steffen Geißler**Certification acc. to:** UL 61010-1:2012 R6.23
CSA C22.2 No. 61010-1-12 (R2022) R6.23**Product Information****Certified Product:** Auto-sampler**Model Designation:** AS vario ER (11-0514-004-26)
AS vario (11-0514-003-26)
AS 21hp (11-0516-003-26)
AS 10e (11-0516-002-26)**Technical Data:** Rated voltage : 24 V DC
Rated current : 2.08 A
Rated power : 50 W
Protection class: III**Remarks:****Appendix:** 1, 1-3**Date of issue:** 2024-12-19
(yr/mo/day)

TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 **TÜVRheinland®**